

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 5005 (1968): Burnisher, Dental, No. Brn 1/2 [MHD 8: Dentistry]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE




IS : 5005 - 1968

Indian Standard

SPECIFICATION FOR
BURNISHER, DENTAL, NO. BRN 1/2

UDC 616.314.72 : 615.472.3



© Copyright 1969

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 1

Price Rs 2.00

Gr 2

April 1969

Indian Standard

SPECIFICATION FOR

BURNISHER, DENTAL, NO. BRN 1/2

Denta 1 Instruments Sectional Committee, CPDC 21

Chairman

COL N. N. BERY

Representing

**Ministry of Health, Family Planning & Urban
Development**

Members

SHRI D. A. BOND

DR RATAN H. DOCTOR

SHRI N. P. DAVER (Alternate)

DR M. S. N. GINWALLA

MAJ-GEN KARTAR SINGH

SHRI L. V. KINI

SHRI V. KRISHNAMOORTHY

Dental Products of India Ltd, Bombay

Indian Dental Traders, Bombay

Nair Hospital Dental College, Bombay

Ministry of Defence (DGAfMS)

Kini Surgical & Engineering Works, Manipal

**Directorate General of Technical Development
(Ministry of Industrial Development & Com-
pany Affairs)**

SHRI V. V. S. MURTHY

**Indian Drugs & Pharmaceuticals Ltd, New
Delhi**

SHRI K. L. TALWAR

Ministry of Defence (DGI)

SHRI A. B. RAO,

Director General, ISI (Ex-officio Member)

Director (Consr Prod)

Secretary

SHRI V. K. KAPOOR

Assistant Director (Consr Prod), ISI

Dental Cutting and Filling Instruments Subcommittee, CPDC 21 : 1

Convener

DR D. C. MIGLANI

Madras Medical College, Madras

Members

SHRI H. AGGARWAL

SHRI D. A. BOND

SHRI L. V. KINI

SHRI J. PRABHAKAR

COL WARDEV SINGH

H. Aggarwal & Bros, Jullundur City

Dental Products of India Ltd, Bombay

Kini Surgical & Engineering Works, Manipal

Indian Drugs & Pharmaceuticals Ltd, New Delhi

Ministry of Defence (DGAfMS)

**INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 1**

Indian Standard

SPECIFICATION FOR BURNISHER, DENTAL, NO. BRN 1/2

0 . F O R E W O R D

0.1 This Indian Standard was adopted by the Indian Standards Institution on 27 December 1968, after the draft finalized by the Dental Instruments Sectional Committee had been approved by the Consumer Products Division Council.

0.2 This standard is one of a series of Indian Standards on dental cutting and filling instruments and has been formulated at the instance of the Advisory Committee for Development of Surgical Instruments, Equipment and Appliances, Government of India. Other standards published so far in the series are given on P 6.

0.3 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS : 2 - 1960*. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

1. SCOPE

1.1 This standard specifies the requirements of dental burnisher No. BRN 1/2.

2. MATERIAL


2.1 The instrument shall be made of stainless steel conforming to Designation 30Cr13 of Schedule V of IS : 1570-1961† or carbon steel conforming to Designation T50 of Schedule VI of IS : 1570 - 1961 †.

3. SHAPE AND DIMENSIONS

3.1 The instrument shall conform to shape and dimensions as given in Fig. 1.

*Rules for rounding off numerical values (revised).

†Schedules for wrought steel for general engineering purposes.



All dimensions in millimetres.

FIG. 1 BURNISHER, DENTAL, NO. BRN 1/2

3.2 Handle — The requirements **prescribed in 3.3 of IS: 3887-1966*** shall apply.

4. WORKMANSHIP AND FINISH

4.1 The working ends of the instrument shall be rounded and shall be free from burrs.

4.2 The instrument made of stainless steel shall be suitably passivated and finished bright all over. Carbon steel instruments shall be plated chromium over nickel and the plating shall conform to Service Grade No. 2 of IS: 1068 -1968†.

5. HEAT TREATMENT

5.1 The stainless steel instrument shall be uniformly hardened and tempered to 370 to 400 *HV*. In case of carbon steel instrument only the working ends shall be hardened and tempered to 400 to 470 *HV*.

6. TESTS

6.1 Practical Test — The working ends of the instrument shall be applied to the top of a compression balance at an angle of 20° approximately. Gripping the handle about the middle of its length, a force of 3 kgf shall be applied for 30 seconds. On release of force, the instrument shall show no sign of damage or a permanent set.

6.2 Test for Corrosion Resistance — The stainless steel instrument shall be tested for corrosion resistance as given in 6.2.1.

6.2.1 Copper Sulphate Test — The instrument shall be scrubbed with the soap and warm water, rinsed in hot water followed by a dip in ethyl alcohol (9.5 percent) and dried. The instrument shall be completely immersed in copper sulphate solution at room temperature for 6 minutes and then washed with fresh water or wiped with wet cotton wool. The copper sulphate solution shall be made up as follows:

Copper sulphate ($\text{CuSO}_4 \cdot 5\text{H}_2\text{O}$)	4.0 g
Sulphuric acid (H_2SO_4) (sp gr 1.84)	10.0 g
Water (H_2O)	90.0 ml

There shall be no red stains or spots on the instrument after the test, but the polished surface becoming dull may be permitted.

*General requirements for cutting type dental instruments.

†Specification for electroplated coatings of nickel and chromium on iron and steel (first revision).

7. MARKING

7.1 The space on the handle shall be marked with the number of instrument, manufacturer's name, initials or trade-mark and the country of manufacture. In case of instruments made of stainless steel, letters, 'SS' shall also be marked on the handle.

7.1.1 Each instrument may also be marked with the ISI Certification Mark.

NOTE — The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act, and the Rules and Regulations made thereunder. Presence of this mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard, under a well-defined system of inspection, testing and quality control during production. This system, which is devised and supervised by ISI and operated by the producer, has the further safeguard that the products so actually marketed are continuously checked by ISI for conformity to the standard. Details of conditions, under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

8. PACKING

8.1 The packing shall conform to 7.1 of IS : 3887-1966*.

*General requirements for cutting type dental instruments.

INDIAN STANDARDS

ON

Dental Cutting and Filling Instruments

IS:	Rs
3887-1966 General requirements for cutting type dental instruments ...	1.50
3888-1967 Spatula (dental)...	2.00
3889-1967 Dental chisels	3.50
3890 (Part I) - 1967 Instruments, plastic filling, dental, Part I Designa- tion No. 1, 3, 4, 5 and 6	3.50
3890 (Part II) - 1967 Instruments, plastic filling, dental, Part II Designa- tion No. 12, 20, 21, 46, 47, 163 and 183	4.00
4689-1968 Enamel hatchet, dental, No. EH 1/2	2.00
4714-1968 Scalers, dental, No. SC 1/2, 3/4 and 5/6	2.50
4715-1968 Excavators, dental, No. EX 1/2, 3/4, 5/6, 7/8, 9/10, 11/12 and 13/14	4.00
4716-1968 Probes, dental, No. PR 1, 2, 3, 4, 5, 6/7 and 8/9	4.00
*5004-1968 Spoons, cutting, dental, No. 1/2, 3/4 and 5/6	
*5017-1968 Carvers, dental, No. CAR 1/2, 3/4 and 5/6	

PUBLICATIONS OF INDIAN STANDARDS INSTITUTION

INDIAN STANDARDS

Over 5 000 Indian Standards, broadly classified under the following main heads, have been issued so far:

Agriculture & Food
Chemical
Civil Engineering
Consumer Products

Electrotechnical
Mechanical Engineering
Structural & Metals
Textile

Of these, the standards belonging to the Consumer Products Group fall under the following categories:

Anaesthesia, Resuscitation & Allied Equipment
Artificial Limbs
Coir
Dental Equipment and Instruments
Domestic Hardware
Fasteners for Consumer Goods
Fountain Pens and Ball Point Pens
Gas Burning Appliances

Hospital Equipment
Hospital Utensils
Medical Glass Instruments
Oil Burning Appliances
Safety Matches
Sports Goods
Surgical Instruments and Dressings
Umbrellas
Utensils and Cutlery

OTHER PUBLICATIONS

	Rs
ISI Bulletin (Published Every Month)	
Single Copy	3-00
Annual Subscription	25-00
Annual Reports (from 1948-49 Onwards)	2-00 to 3-00 each
Handbook of ISI Publications, 1968	5-00

Available from

INDIAN STANDARDS INSTITUTION

Headquarters

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 1

Telephones 27 36 11 - 20

Telegrams Manaksanstha

Branch Offices

Telegrams Manaksanstha

534 Sardar Vallabhbhai Patel Road
5 Chowringhee Approach
5-9-201/2 Chirag Ali Lane
117/418 B Sarvodaya Nagar
54 General Patters Road

Bombay 7	Telephone	35 69 44
Calcutta 13	"	23-18 23
Hyderabad 1	"	3 23 35
Kanpur 5	"	6 76 95
Madras 2	"	8 72 78